



Attestation of Conformity

No. ICR Polska/M4400137



**Name and address
of Registered Manufacturer:**

ZHEJIANG MONDIAL ELECTRONIC TECHNOLOGY CO.,LTD
NO.15 CHAOYANG ROAD, HAIYOU SUBDISTRICT, SANMEN
COUNTY, TAIZHOU CITY, 317100, ZHEJIANG

Product name:

Infrared Forehead Thermometer

Product type/model:

MD-33520

Trade mark:

n/a

This Attestation confirms that the product meets the requirements of the following normative documents and within limits of its documents gives presumption of conformity with essential requirements of Directive 93/42/EEC.

Relevant EC Directive:

Medical Device Directive 93/42/EEC

**Conformity assessment
procedure:**

EC Declaration of Conformity (Annex VII of Directive
93/42/EEC)

Classification:

Class I according Rule 1 of Annex IX of Directive 93/42/EEC

Applied normative documents:

EN 60601-1: 2006
EN 60601-1-2: 2015

**Applied Quality
Management System**

n/a

This AoC will remain valid only if Quality Management System Certificate remains valid.
The assessment process has been carried out in accordance with the program PC-P-07-07.
Evaluation has been carried out in accordance with test report made by:

- Shenzhen Labman Testing Technology Co., Ltd.

No. of test report:

LM-202003060R

Issue date:

25.03.2020

Expiration date:

24.03.2025

The mutual obligations and rights of the certification are regulated by the contract No. ICR
Polska/2020-3125.

This Attestation applies to products having the same attributes (parameters), intended use, that have
been evaluated and meet the requirements of the aforementioned standard.



Director: Rafał Kalinowski

Warsaw, 25. 03. 2020.

ICR Polska Co. Ltd.

ul. Plac Przymierza 6, 03-944 Warszawa
www.icrpolska.com, e-mail: icrpolska@icrq.com

